

Head Office

44-1, 2-chome, Hatagaya, Shibuya-ku, Tokyo 151-0072, Japan
Telephone: 81-3-3374-8111 Facsimile: 81-3-3374-8399

Rev.4

DECLARATION of CONFORMITY

We, TERUMO Corporation
44-1, 2-chome, Hatagaya, Shibuya-ku, Tokyo 151-0072 JAPAN

declare under our sole responsibility that the product:

Product : Accessories related to Infusion Pumps and Syringe Infusion Pumps

Model/Type : TERUFUSION Rack System

to which this declaration relates is in conformity with the essential requirements and other relevant requirements of the Radio Equipment Directive (2014/53/EU).
The product is in conformity with the following standards:

RADIO : EN 300 328 V2.1.1:2016-11

**EMC : EN 301 489-1 V1.9.2:2011-09
EN 301 489-17 V2.2.1:2012-09
EN 60601-1-2:2001/A1:2006
EN 60601-2-24:1998**

**SAFETY : EN 60950-1:2006+Amd.11:2009+Amd.1:2010+Amd.12:2011+Amd.2:2013
EN 62479:2010
EN 60601-1:1990/A1:1993/AC:1994/A2:1995
EN 60601-1-1:2001
EN 60601-1-2:2001/A1:2006
EN 60601-1-4:1996/A1:1999
IEC 60601-1-4:2000
EN 60601-1-6:2004
EN 60601-1-6:2007
EN 60601-2-24:1998**

Conformity test for radio part is performed with Wireless LAN Module (Model: UGFZ1)



Date : June 1, 2017

Signature : 

Name : Toshio Nakashima

Title : General Manager, Quality Assurance Department



Head Office

44-1, 2-chome, Hatagaya, Shibuya-ku, Tokyo 151-0072, Japan
Telephone: 81-3-3374-8111 Facsimile: 81-3-3374-8399

Rev.4

Object of Declaration

T E * R S □ □ □ □ □ □
1 2 3 4 5 6 7 8 9 10 11

Digits	Display items	Display
1-2	Product name	TE:TERUFUSION ELECTRICITY
3	Destination	* :For export
4-5	Type	RS : Rack System
6-11	Product model	800N :Communication rack system 800N03:Communication rack system (TE only)