



EU DECLARATION OF CONFORMITY

We, TERUMO CORPORATION
44-1, 2-chome, Hatagaya, Shibuya-ku, Tokyo 151-0072, Japan
with Single Registration Number: JP-MF-000017478

being the manufacturer of:

TERUFUSION Rack System

[INFUSION INSTRUMENTS - HARDWARE]

Intended purpose:

TERUFUSION Standard Rack System

The TERUFUSION Standard Rack System, which is intended to be used by healthcare professionals, is a modular system designed to meet the request of a multi-infusion system which combines the advantages of a stacking system (compact and extendable) with those of a racking system (the pumps can be removed or added independently). This modular system allows the use of multiple TERUFUSION Syringe Pump and TERUFUSION Infusion Pump for the same patient with only one power cable.

TERUFUSION Communication Rack System

The TERUFUSION Communication Rack System, which is intended to be used by healthcare professionals, is designed for racking multiple pumps and, provides one-step attaching and detaching of the specified syringe pump and infusion pump, and supplies the AC source to the pump attached. In addition, this product mediates the communication between the pumps or the pump and an external device.

Basic UDI-DI: 498735028RS3S

Related product codes: See Appendix A (full list of active codes)

declare that the above product of **Class I** is in conformity with the applicable requirements of the Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices, and has been subject to the conformity assessment procedure laid down in Article 52.7 of the Regulation, has been drawn up.

There is no reference to Common Specifications that have been used to within the conformity assessment for Regulation (EU) 2017/745.

We furthermore declare that the above product is in conformity with the provisions of RoHS Directive (2011/65/EU as amended and applicable) and RED (2014/53/EU) : See Appendix B.



Rev. 04
DoC No. DOC- ME-TDP0680A
Reference to. ME-TDP0680A

Authorised Representative: TERUMO EUROPE N.V.
Authorised Address: Interleuvenlaan 40, 3001 Leuven, Belgium
with Single Registration Number: BE-AR-000001433

This EU declaration of conformity is issued under our sole responsibility.

Tokyo, 2025-06-17

(place and date of issue)

署名者:
hiroki sasagawa
署名者名: hiroki sasagawa
署名理由: この文書を承認する
署名時刻: 2025-06-17 | 1:43:25 午後 JST
3F7B1A275BA74182B04CDC2E69B7202D

Hiroki Sasagawa
General Manager
Quality Assurance Department
For and on behalf of
TERUMO CORPORATION



Appendix A – Related product codes

< Regulation (EU) 2017/745 and RoHS Directive >

Product code									UDI-DI code															
T	E	*	R	S	7	0	0	N	0	4	9	8	7	3	5	0	7	1	1	7	8	6		
T	E	*	R	S	8	0	0	N	0	4	9	8	7	3	5	0	7	1	1	8	0	9		
T	E	*	R	S	8	1	1	N	0	4	9	8	7	3	5	0	7	1	1	8	2	3		
Product code									UDI-DI code															
T	E	*	R	S	8	0	0	N	0	3	0	4	9	8	7	3	5	0	7	1	2	3	0	1

< RED >

Product code									UDI-DI code															
T	E	*	R	S	8	0	0	N	0	4	9	8	7	3	5	0	7	1	1	8	0	9		
Product code									UDI-DI code															
T	E	*	R	S	8	0	0	N	0	3	0	4	9	8	7	3	5	0	7	1	2	3	0	1

Appendix B – Reference standard of RoHS and RED

< RoHS Directive >

EN IEC 63000:2018

< RED >

RADIO

EN 300 328 V2.2.2:2019-07

EMC

EN 301 489-1 V2.2.3:2019-11

EN 301 489-17 V3.2.4:2020-09

EN 60601-1-2:2015+A1:2021

EN 60601-2-24:2015

SAFETY

EN 60601-1:2006+A1:2013+A12:2014+A2:2021

EN 60601-1-2:2015+A1:2021

EN 60601-1-6:2010+A1:2015+A2:2021

EN 60601-2-24:2015

EN 62366-1:2015+A1:2020

EN IEC 62368-1:2020+A11:2020

EN 62479:2010

Conformity test for radio part is performed with Wireless LAN Module (Model:UGFZ1)